



中国国家强制性产品认证证书



证书编号: 2020322304003143

认证委托人名称: WIKA Instruments Italia Srl
认证委托人地址: Via Guglielmo Marconi, 8 20044 Arese (MI) Italy
生产者名称: WIKA Instruments Italia Srl
生产者地址: Via Guglielmo Marconi, 8 20044 Arese (MI) Italy
生产企业名称: WIKA Instruments Italia Srl
生产企业地址: Viale DeGasperi, 48 20008 Bareggio (MI) Italy

产品名称: 波登管式压力开关

系列、规格、型号: BAX 系列

标准: GB 3836.1-2010、GB 3836.2-2010、GB 12476.1-2013、
GB 12476.5-2013

上述产品符合强制性产品认证实施规则 CNCA-C23-01:2019 的要求, 特发此证。

发证日期: 2020 年 11 月 4 日

有效期至: 2025 年 11 月 3 日

首次发证日期: 2020 年 11 月 4 日

证书有效期内本证书的有效性依据发证机构的定期监督获得保持。

本证书的相关信息可通过国家认监委网站 www.cnca.gov.cn 查询



批准:



上海仪器仪表自控系统检验测试所有限公司



CERTIFICATE FOR CHINA COMPULSORY PRODUCT CERTIFICATION



CERTIFICATE NO: 2020322304003143

APPLICANT: WIKA Instruments Italia Srl
ADDRESS: Via Guglielmo Marconi, 8 20044 Arese (MI) Italy
MANUFACTURER: WIKA Instruments Italia Srl
ADDRESS: Via Guglielmo Marconi, 8 20044 Arese (MI) Italy
FACTORY: WIKA Instruments Italia Srl
ADDRESS: Viale DeGasperi, 48 20008 Bareggio (MI) Italy
PRODUCTNAME: Bourdon pressure switches
SERIES,SPECIFICATION,MODEL: BAX Series
STANDARDS: GB 3836.1-2010、GB 3836.2-2010、GB 12476.1-2013、
GB 12476.5-2013

This is to certify that the above mentioned product(s) complies with the requirements of implementation rules for compulsory certification (REFNO. CNCA-C23-01:2019).

Valid from: November 4, 2020

Valid until: November 3, 2025

Date of original certification: November 4, 2020

The validity of this certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.

This certificate is available through CNCA's website: www.cnca.gov.cn



APPROVAL:

Xu JianPing



Shanghai Inspection and Testing Institute of Instruments and Automation Systems Co., Ltd.

<http://www.sitiilas.com.cn>

Building 9, 103 Cao Bao Road, Shanghai 200233, China

Tel: +86 21 64510844

S 0002114



中国国家强制性产品认证证书



证书编号: 2020322304003143

附 件

产品名称: 波登管式压力开关

型号规格: BAX[a] - *...
[a] 表示电接点 (数量+类型) 代号, 可为 U (N, S, O, G, R) 或 D (N, S, O, G, R)。
- *... 后参数表示工作参数和连接螺纹类型 (与防爆性能无关)。

防爆标志: Ex d IIC T4/T6 Ga/Gb
Ex tD A21 IP66 T85°C/T135°C

电气参数:

电接点代码 [a]	触电容量	
	CA/AC	CC/DC
UN、DN	15A 250V	2A 24V 0.5A 125V 0.25A 220V
US、DS	15A 250V	2A 24V 0.5A 220V
UO、DO	1A 125V	0.5A 24V
UG、DG	1A 125V	0.5A 24V
UR	20A 250V	2A 24V 0.5A 220V
DR	15A 250V	2A 24V 0.5A 220V

批 准:



上海仪器仪表自控系统检验测试所有限公司