

[1]

EU-TYPE EXAMINATION CERTIFICATE

[2] Directive 2014/34/EU of the European Parliament and of the Council of 26 February 2014

[3] EU-Type Examination Certificate Number: **Presafe 19 ATEX 21636X** **Issue 2**

[4] Product: **Differential and absolute pressure switches series DA and APA**

[5] Manufacturer: **Euromisure WIKA Instruments S.a.s. di WIKA Italia S.r.l.**

[6] Address: **Viale De Gasperi, 48
20008 Bareggio (MI) - ITALY**

[7] This product and any acceptable variation thereto is specified in the schedule to this certificate and the documents therein referred to.

[8] DNV Product Assurance AS, notified body number 2460, in accordance with Article 17 of Directive 2014/34/EU of the European Parliament and of the Council, dated 26 February 2014, certifies that this product has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of products intended for use in potentially explosive atmospheres, given in Annex II to the Directive.

The examination and test results are recorded in confidential reports listed in item 16.

[9] Compliance with the Essential Health and Safety Requirements has been assured by compliance with: **EN IEC 60079-0:2018 and EN 60079-1:2014; EN 60079-26:2015 and EN 60079-31:2014**

Where additional criteria beyond those given here have been used, they are listed at item 18 in the Schedule.

[10] If the sign "X" is placed after the certificate number, it indicates that the product is subject to the "Specific Conditions of Use" listed under item 17 of this certificate.

[11] This EU-TYPE EXAMINATION CERTIFICATE relates only to the technical design of the specified product in accordance to the Directive 2014/34/EU. Further requirements of the Directive apply to the manufacturing process and supply of this product. These are not covered by this certificate.

[12] The marking of the product shall include the following:

 **II 1/2 G Ex db IIC T6 or T4 Ga/Gb**

 **II 1/2 D Ex tb IIIC T85°C or T135°C Db**



Date of issue:
2025-04-07



Asle Kaastad
For DNV Product Assurance AS
The Certificate has been digitally signed.

[13]

Schedule

[14] **EU-Type Examination Certificate No:**

Presafe 19 ATEX 21636X

Issue 2

[15] **Description of Product**

The CCE02 enclosure for differential and absolute pressure switches series DA**** and APA**** is intended to contain one or two microswitches (contacts) and a terminal block used to electrically connect them to an external control circuit. The microswitches are operated by a rotation of a flameproof joint (operating rod) connected by a differential (absolute) pressure cell. The degree of protection is IP66 according to the Standard IEC 60529:2001

Type designation

Pressure switches series DA and APA (CCE02)

The instruments series DA**** and APA**** are classified as electrical equipment for potentially explosive atmospheres because of the presence of: gas, vapors and fogs of the IIC group, combustible dusts of the IIIC group.

The instruments of the series DA*** and APA**** are 1/2 category equipment intended to be used on surface (not in mines).

Electrical Data

380 V / 20 A c.a. or 250V / 20 A c.a.
220 V / 0.5 A (30V / 2A) d.c.

Degrees of protection (IP Code)

IP66

Ambient temperature:

-60°C to +60°C for T6
-60°C to +85°C for T4

Pressure switch: process temperature (media temperature at the process connection)

The limits of process temperature are: -60 ÷ +85 °C

Temperature Classification	Ambient temperature range (Tamb)	Min process temperature (Media temperature at the process connection of the instrument)	Max process temperature (Media temperature at the process connection of the instrument)	Max Current (resistive load)
T6	-60 .. +60 °C	-60 °C	+60 °C	(depending on the microswitches selected) See Nameplate
T5				
T4	-60 .. +85 °C			
T3				
T2				
T1				

Warning label

Power off before opening.

If the ambient temperature exceeds 60°C, use suitable cable for operating temperature > 105°C

Installation conditions

The accessories used for cable entries and for closing unused openings shall be certified according to EN 60079-0, EN 60079-1 and EN 60079-31 standards.

A minimum degree of protection IP66 shall be guaranteed according to EN 60529 standard.

Routine tests

The manufacturer shall carried out the routine test prescribed at clauses 27 of the EN 60079-0 standard.

The equipment are exempted from overpressure routine test, according to clause 16.2 of the EN 60079-1 standard, since they have been submitted to an overpressure test at 4 times reference pressure under 40.8 bar for ambient temperature of –60°C.

[16] **Report No.:** SC187565-20200701-DA-TLM-21636_02
Project No: PRJN-767532

[17] Specific Condition(s) of Use

1. The instruments series DA**** and APA**** can be mounted on a boundary wall between zone 20 and zone 21.
2. Due to the thickness less than 1mm of the membrane, the user should take into consideration that the equipment shall not be subject to environmental conditions which might adversely affect the partition membrane.
3. The width of the flameproof joints is greater than the values specified in the tables of 60079-1 standard.

[18] Essential Health and Safety Requirements

Met by compliance with the requirements mentioned in item 9.

[19] Drawings and documents

Number	Title	Rev.	Date
SP-TC.791E	Technical specification	5	2024-06
NI.709E	Safety Instructions	3	2024-06
CCE02U	Drawing CCE02	2	2024/06/06
CCE02D	Drawing CCE02	4	2024/06/06
SP-TC.787E	Data sheet sealing material	1	2024-06
DGT.ATEX-A	Drawing marking plate	3	2024/06/17
DGT.LABEL-A	Drawing warning plate	4	2024/06/17

[20] **Certificate History**

Issue	Description	Issue date	Report no.
0	Original issue	2020-03-25	SC166368-20191202-DA-TLM-21636
1	Change of company name	2020-07-28	SC187565-20200701-DA-TLM-21636
2	Change of company name	2025-04-07	SC187565-20200701-DA-TLM-21636 02

Compliance of the product with the applicable safety requirements of the relevant industrial standards has not been verified and is not covered by this certificate.

END OF CERTIFICATE

