



Certificate of Compliance

Certificate: 80224078 **Master Contract:** 166720
Project: 80224078 **Date Issued:** 2025-04-10
Issued to: WIKA Alexander Wiegand SE & Co. KG
Alexander-Wiegand-Str 30
Klingenberg, Bavaria 63911
Germany
Issued by: Cem Kayadibi
Cem Kayadibi
Attention: Jonathan Vion

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



PRODUCTS

Class 2252 06 PROCESS CONTROL EQUIPMENT - Process Control Equipment
Class 2252 86 PROCESS CONTROL EQUIPMENT - Certified to US Standards

Telemetry Unit, Type: NETRIS®3 *L2*

Equipment powered by a single 3.0VDC (size “A”, Li MNO2, 3.3Ah) cell.

Model(s)	Battery (VDC)	Ambient Temp. Range (°C)	Enclosure Type Rating
NETRIS®3 - *L2*	3	-40 to +60	4X



Certificate: 80224078

Master Contract: 166720

Project: 80224078

Date Issued: 2025-04-10

Notes:

1. The above model is battery operated: 3.0V DC (size "A", Li-MNO₂, 3.3Ah), for fixed installation and permanently connected, Pollution Degree 4, Type 4X Enclosure.
2. Mode of operation: Continuous.
3. Environmental Conditions: (Extended) Ambient temperature range: -40°C to + 60°C, Indoor and Outdoor Use, Relative Humidity: 20...90%, non-condensing, Altitude: Max. 2000m.
4. Maximum defined output parameters:
Output Parameters:
 $U_O \leq 5.88V$
 $I_O \leq 200mA$
 $P_O \leq 295mW$

Description of model differences:

The stars can be replaced by the following but do not affect the design of the equipment.

NETRIS®3 – *L2*

- First *: Certification marking (the star can be replaced by any number ≥ 1)
- Second *: On-site mounting interface (the star can be replaced by any letter from A to Z)

APPLICABLE REQUIREMENTS

CSA C22.2 No. 61010-1-12, UPD1:2015, UPD2:2016, AMD1:2018, UPD3:2023 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements - Third Edition; Update No. 1: July 2015; Update No. 2: April 2016; Update No. 3: June 2023

UL 61010-1 3rd ed (Rev. Jun 6, 2023) - UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements

CSA C22.2 No. 94.2:20 - Third Edition - Enclosures for electrical equipment, environmental considerations

ANSI/UL 50E, 3rd Edition - Enclosures for Electrical Equipment, Non-Environmental Considerations

Conditions Of Acceptability

N/A - No Conditions of Acceptability.

Markings

See Report.



Certificate: 80224078


Project: 80224078

Master Contract: 166720

Date Issued: 2025-04-10

Notes:

Products certified under Class(es) C225206, C225286 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80224078

Master Contract: 166720

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80224078	2025-04-10	Original cCSAus certification of Telemetry Unit model NETRIS®3 – *L2* based on CB report 80224082, including evaluation of enclosure estimated for Type 4X according to CSA C22.2 No. 94.2 and UL 50E.