



Certificate of Compliance

Certificate: 2706069

Master Contract: 166720

Project: 2706069

Date Issued: November 4, 2014

Issued to: WIKA Alexander Wiegand SE & Co. KG

Alexander-Wiegand-Str 30
Klingenberg, 63911
Germany
Attention: Simon Edler

The products listed below are eligible to bear the CSA Mark shown



Alan McLaughlin

Issued by: Alan McLaughlin

PRODUCTS

CLASS 2252 06 - PROCESS CONTROL EQUIPMENT

Pressure Transmitters, Models S-10 and S-11, rated DC +30 V maximum (Class 2 supply source), Enclosure Type IP65/IP67/IP68 or 4/4X, maximum pressure 4000 bar (400 MPa) or 58000 psi, Optional: Single Seal, Ambient temperature range -20 °C to +85 °C.

Pressure Transmitters, Models C-10, SA-11, LS-10 and LH-10, rated DC +30 V maximum (Class 2 supply source), Enclosure Type IP65/IP67/IP68 or 4/4X, maximum pressure 4000 bar (400 MPa) or 58000 psi, Ambient temperature range -20 °C to +85 °C as applicable.

Notes:

- Models designation maybe followed by up to 16 numbers and letters, maybe separated by “-“
- The above model is Equipment Class III, Pollution Degree 2, Installation Category I.
- Mode of operation: Continuous
- Environmental Conditions: Normal: -20 to 85 °C, 2000 m max, 80% to temperatures up to 31 °C decreasing linearly to 50% rH at 40 °C;

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements