

Herstelleraussage

14150645.05

Verordnung (EG) Nr. 1935/2004

Wir bestätigen hiermit für die Materialien und Gegenstände, die bei bestimmungsgemäßem Gebrauch in Kontakt mit Lebensmitteln kommen können, die Konformität der metallischen messstoffberührten Materialien mit den allgemeinen Anforderungen der

**Verordnung (EG) Nr. 1935/2004
vom 27. Oktober 2004
über Materialien und Gegenstände, die dazu
bestimmt sind, mit Lebensmitteln in
Berührung zu kommen.**

Dies gilt für folgende Typen:

**Druckmessgerät mit frontbündiger
Plattenfeder für die sterile Verfahrenstechnik**

**PG43SA-C
PG43SA-D
PG43SA-S**

Alle messstoffberührten Teile sind aus Edelstahl:

316L/1.4435, Inconel 718/2.4668

Die Werkstoffeigenschaften werden bei Bestellung mit Materialzertifikat 3.1 nach EN 10204 detailliert nachgewiesen.

Die Eignung zum Einsatz, sowie die geeignete Reinigung und Wartung, ist bezogen auf die Applikation vom Anwender sicher zustellen.

Die Fertigungsprozesse der Geräte wurde nach "Guter Herstellungspraxis", wie in der Verordnung (EG) Nr. 2023/2006 „über gute Herstellungspraxis für Materialien und Gegenstände, die dazu bestimmt sind, mit Lebensmitteln in Berührung zu kommen“ vom 22. Dezember 2006 beschrieben, durchgeführt.

Unterzeichnet für und im Namen von / Signed for and on behalf of

WIKA Alexander Wiegand SE & Co. KG

Klingenberg, 2020-07-06

Company Division Pressure

i.V.



Alfred Häfner

Unterschrift autorisiert durch das Unternehmen / Signature authorized by the company

WIKA Alexander Wiegand SE & Co. KG
Alexander-Wiegand-Straße 30
63911 Klingenberg
Germany

Tel +49 9372 132-0
Fax +49 9372 132-406
E-Mail info@wika.de
www.wika.de

Kommanditgesellschaft: Sitz Klingenberg -
Amtsgericht Aschaffenburg HRA 1819
Komplementärin: WIKA Verwaltungs SE & Co. KG -
Sitz Klingenberg - Amtsgericht Aschaffenburg
HRA 4685

Manufacturer Statement

14150645.05

Regulation (EC) Nr. 1935/2004

For the materials and articles which may come into contact with food when used as intended, we hereby confirm the conformity of the wetted metal materials with the general requirements of

**Regulation (EC) No. 1935/2004
of 27 October 2004
on materials and articles intended to come
into contact with food.**

This applies to the following models:

**Flush diaphragm pressure gauge
for sanitary applications**

**PG43SA-C
PG43SA-D
PG43SA-S**

All wetted parts are made of stainless steel:

316L/1.4435, Inconel 718/2.4668

On order with 3.1 material certificate the material characteristics are verified in detail according to EN 10204.

Based on the specific application, the suitability for use as well as the proper cleaning and maintenance has to be ensured by the user.

The manufacturing process has followed „Good Manufacturing Practice (GMP)“ as specified in the regulation (EC) No 2023/2006 “on good manufacturing practice for materials and articles intended to come into contact with food” of 22. December 2006.

Quality Management Pressure

i.A.



Roland Stapf

Komplementärin:
WIKA International SE - Sitz Klingenberg -
Amtsgericht Aschaffenburg HRB 10505
Vorstand: Alexander Wiegand
Vorsitzender des Aufsichtsrats: Dr. Max Egli