

UK Declaration of Conformity

Document No. 14545221

Issue 01

We declare under our sole responsibility that the UKCA marked products

Type Designation	CPT6020
Description	Precision Pressure Transducer, Basic Version
according to the valid data sheet	CT 25.13
comply with the applicable UK Statutory Requirements	Applied designated standards ⁽¹⁾
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 No. 3032 (as amended)	EN IEC 63000:2018
Electromagnetic Compatibility Regulations 2016 No. 1091 (as amended)	EN 61326-1:2013 stimmt auch überein mit/ <i>also complies with</i> EN IEC 61326-1:2021
Pressure Equipment (Safety) Regulations 2016 No. 1105 (as amended) ⁽²⁾	[-]

- (1) At the time of writing this document designated standards are prefixed "BS", "EN", "EN ISO" or "EN IEC". Where the designated standard specified in the notice of publication is prefixed "EN" it is acceptable to reference this version in technical documentation, or a version of the same standard with a national prefix. For more information see <https://www.gov.uk/guidance/designated-standards>.
- (2) PS > 200 bar; Module A, pressure accessory

Signed for and on behalf of

Mensor LP

201 Barnes Dr., San Marcos. Texas 78666, USA

2026-06-08



Jason Otto, Vice President
Calibration Technology



Doyle Butler, Director of Quality
Calibration Technology