

Differential pressure sensor / Differential pressure air flow sensor / Differential pressure air flow controller, models A2G-500, A2G-520 and A2G-540

EN



Differential pressure sensor, model A2G-500
Differential pressure air flow sensor, model A2G-520
Differential pressure air flow controller, model A2G-540



© 07/2024 WIKA Alexander Wiegand SE & Co. KG
All rights reserved.
WIKA® is a registered trademark in various countries.

Prior to starting any work, read the operating instructions.
Keep for later use.

Contents

1. General information	4
1.1 Abbreviations, definitions	4
1.2 Explanation of symbols	5
2. Safety	5
2.1 Intended use	5
2.2 Improper use	6
2.3 Personnel qualification	6
2.4 Labelling, safety markings	7
2.5 Contents of the QR code	8
3. Transport, packaging and storage	9
3.1 Transport.	9
3.2 Packaging and storage	9
4. Design and function	10
4.1 Overview.	10
4.2 Scope of delivery	10
4.3 Description	11
4.4 Overview of versions.	12
5. Commissioning and operation	13
5.1 Commissioning procedure	13
5.2 Instrument mounting	14
5.3 Electrical connection.	15
5.4 Operation / Instrument setting.	20
6. Faults	44
7. Maintenance and cleaning	47
7.1 Maintenance	47
7.2 Changing the batteries (only with instruments with battery operation)	47
7.3 Cleaning	47
8. Dismounting, return and disposal	48
8.1 Dismounting	48
8.2 Return.	48
8.3 Disposal	49
9. Specifications	51
9.1 Specifications A2G-5x0.	51
9.2 Approvals	56
9.3 Certificates	56
9.4 Dimensions in mm [in]	57
10. Accessories and spare parts	58
Annex: EU declaration of conformity	59

1. General information

- The instrument described in the operating instructions has been designed and manufactured using state-of-the-art technology. All components are subject to stringent quality and environmental criteria during production. Our management systems are certified in accordance with ISO 9001 and ISO 14001.
- These operating instructions contain important information on handling the instrument. Working safely requires that all safety instructions and work instructions are observed.
- Observe the relevant local accident prevention regulations and general safety regulations for the instrument's range of use.
- The operating instructions are part of the product and must be kept in the immediate vicinity of the instrument and readily accessible to skilled personnel at any time. Pass the operating instructions on to the next operator or owner of the instrument.
- Skilled personnel must have carefully read and understood the operating instructions prior to beginning any work.
- In case of a different interpretation of the translated and the English operating instructions, the English wording shall prevail.
- If available, the provided supplier documentation is also considered to be part of the product in addition to these operating instructions.
- The general terms and conditions contained in the sales documentation shall apply.
- Subject to technical modifications.
- Further information:
 - Internet address: www.wika.de / www.wika.com
 - Relevant data sheets: PE 88.05, PE 88.06, PE 88.07
 - Special documentation: Special documentation for LoRaWAN®
 - Contact: info@wika.ch

1.1 Abbreviations, definitions

- Bullet
- Instruction
- 1. ... x. Follow the instruction step by step
- ⇒ Result of an instruction
- See ... cross-references

1.2 Explanation of symbols



DANGER!

... indicates a directly dangerous situation resulting in serious injury or death, if not avoided.



WARNING!

... indicates a potentially dangerous situation that can result in serious injury or death, if not avoided.



CAUTION!

... indicates a potentially dangerous situation that can result in light injuries or damage to property or the environment, if not avoided.



Information

... points out useful tips, recommendations and information for efficient and trouble-free operation.

2. Safety

2.1 Intended use

This differential pressure sensor is used for measuring differential pressure, over-pressure and vacuum of air and other non-inflammable and non-aggressive gases in ventilation and air-conditioning applications and is used in industrial electromagnetic environments.

This instrument is not permitted to be used in hazardous areas.

The instrument has been designed and engineered solely for the intended use described here, and may only be used accordingly.

The technical specifications contained in these operating instructions, see chapter 9 "Specifications", must be observed. It is assumed that the instrument is handled properly and within its technical specifications.

The manufacturer shall not be liable for claims of any type based on operation contrary to the intended use.

2. Safety

2.2 Improper use

- Any use beyond or different to the intended use is considered as improper use.
- Unauthorised modifications to the instrument are not permissible.

2.3 Personnel qualification



The activities described in these operating instructions may only be carried out by skilled personnel who have the qualifications described below.

Skilled electrical personnel


Skilled electrical personnel are understood to be personnel who, based on their technical training, know-how and experience as well as their knowledge of country-specific regulations, current standards and directives, are capable of carrying out work on electrical systems and independently recognising and avoiding potential hazards. The skilled electrical personnel have been specifically trained for the work environment they are working in and know the relevant standards and regulations. The skilled electrical personnel must comply with current legal accident prevention regulations.

2. Safety

2.4 Labelling, safety markings

The labelling, safety markings must be maintained in a legible condition.

Product label (example)



- ① Model
- ② Item number
- ③ Serial number
- ④ QR code per LoRa Alliance® specification
- ⑤ Radio standard
- ⑥ Date of manufacture (YYYY/MM)
- ⑦ Approval marking
- ⑧ Output signal / Bus protocol (optional)
- ⑨ Auxiliary power
- ⑩ Operating range



Do not dispose of with household waste. Ensure a proper disposal in accordance with national regulations.



Before mounting and commissioning the instrument, ensure you read the operating instructions.

2. Safety



In electrical engineering, the protection class is used to categorise and label electrical equipment with regard to the protective and safety measures in place to prevent electric shock.



NFC means “Near Field Communication”. This refers to contactless data transmission that utilises radio frequency identification (RFID) technology.



IP ingress protection ratings indicate the degree to which an object is protected against the ingress of water and dust or other possible external influences.

2.5 Contents of the QR code

Example:

LW (LoRaWAN®):

D0:70B3D597B0000008:70B3D597B0004D71:02A30008:S2Y01FZFCOJ6:CC9B3

Contents	Identifier	Example	Comment
SchemaID	-	D0	-
JoinEUI (64 bit)	-	70B3D597B0000008	JoinEUI = AppEUI
DevEUI (64 bit)	-	70B3D597B0004D71	-
ProfileID	-	02A30008	02A3 = VendorID WIKA 0008 last 4 digits of the JoinEUI
SerNum	S	2Y01FZFCOJ6	Alphanumeric WIKA serial number (11-digit)
Checksum	C	C9B3	-

3. Transport, packaging and storage

3. Transport, packaging and storage

3.1 Transport



WARNING!

Damage from batteries through improper transport

If loose or removed batteries are transported incorrectly, they can explode, burn or leak.

- ▶ Tape exposed contacts and pack the batteries so that they do not move in the packaging (prevent short-circuit).
- ▶ Be careful when transporting.



CAUTION!

Damage through improper transport

With improper transport, damage to property can occur.

- ▶ When unloading packed goods upon delivery as well as during internal transport, proceed carefully.
- ▶ With internal transport, observe the instructions in chapter 3.2 "Packaging and storage".

Check the instrument for any damage that may have been caused.

In the event of any damage, do not commission the instrument and contact the manufacturer immediately.

If the instrument is transported from a cold into a warm environment, the formation of condensation may result in instrument malfunction. Prior to recommissioning, wait for the instrument temperature and the room temperature to equalise.

3.2 Packaging and storage

Do not remove packaging until just before mounting.


Keep the packaging as it will provide optimum protection during transport (e.g. change in place of use, sending for repair).

EN

4. Design and function

4. Design and function

4.1 Overview



- | | |
|--|---|
| ① Display (option) | ⑧ Process connection $\ominus$ |
| ② Cover with insert sheet | ⑨ LoRaWAN® (option) |
| ③ Mounting plate | ⑩ NFC antenna |
| ④ Case | ⑪ Relay (option) |
| ⑤ M20 cable gland (depending on version) | ⑫ Connection terminals |
| ⑥ Blind plug (depending on version) | ⑬ Valve for automatic zero point setting (option) |
| ⑦ Process connection $\oplus$ | |



The illustrations are symbolic images and there may be differences depending on the version.

4.2 Scope of delivery

- Differential pressure sensor
- Mounting plate
- 4 pan-head, self-tapping screws ST4.2x9.5 (cross head Philipps Form H2)
- LoRaWAN® login credentials for commissioning (only with LoRaWAN® devices)
- 2 batteries AA 3.6 V, 2.6 Ah (only with battery version)
- Quick start instructions

Cross-check scope of delivery with delivery note.

4.3 Description

The A2G-500 differential pressure sensor is used to measure differential pressure, gauge pressure and vacuum. The A2G-520 differential pressure air flow sensor measures the pressure difference on components such as ventilators and pitot tubes or similar, e.g. model A2G-FM, and calculates the air flow using the calibration factor (K value). The A2G-540 differential pressure air flow sensor is further used to control differential pressure and air flow.

All instrument versions are used in air and also in non-aggressive and non-inflammable gases in ventilation and air-conditioning systems.

The measured values are available as analogue voltage and current signals, digitally via the RS-485 interface using Modbus[®] RTU or via wireless transmission via LoRaWAN[®]. LoRaWAN[®] is based on LPWAN technology ("Low Power Wide Area Network") to enable high transmission ranges. The IIoT-capable instrument fulfils safety-related requirements of the relevant standards and regulations for on-site display, as well as the requirements of the Radio Equipment Directive for data communication. Integration into any control system or directly into cloud solutions is therefore easily possible.

The settings, measured values and the instrument status can be set and queried on site using a mobile device, such as a smartphone, via the WIKA app "myWIKa wireless device" (NFC) and, depending on the version, by means of buttons or LED operation.

4. Design and function

4.4 Overview of versions

Model	A2G-500	A2G-500 /-520 / -540		A2G-500
Specification	Single-range	Multi-range	Modbus RTU	Wireless (battery power)
Differential pressure	✓	✓	✓	✓
Air flow ¹⁾	✗	O	O	✗
Controller ²⁾	✗	O	O	✗
Operating range				
±250 Pa	✓	✓	✓	✓
±2,500 Pa	✓	✓	✓	✓
±7,000 Pa	✓	✓	✓	✓
±12,000 Pa ³⁾	✓	✓	✓	✓
Measuring range				
Preset	1	8	8	8
Customer-specific	✗	✓	✓	✓
1. Analogue output (only one selection possible)				
2-wire 4 ... 20 mA	O	✗	✗	✗
3-wire 4 ... 20 mA, 0 ... 10 V, 0 ... 5 V or 2 ... 10 V	O	O	O	✗
2. Analogue output				
3-wire 4 ... 20 mA, 0 ... 10 V, 0 ... 5 V or 2 ... 10 V	✗	O	O	✗
Signal inputs	✗	O	O	✗
Relay	✗	O	O	✗
Manual zero point setting	✓	✓	✓	✓
Automatic zero point setting	✗	O	O	✗
Wireless signal (LoRaWAN [®])	✗	O	O	✓
Display	✗	O	O	✗

✓: included

✗: not included

O: optional

1) A2G-520 or A2G-540

2) A2G-540

3) This operating range is only available for A2G-500 and A2G-540 (differential pressure sensor).

40452360.02 04/2025 EN

5. Commissioning and operation

Personnel: skilled electrical personnel

Tools: voltage tester, screwdriver

Only use original parts, see chapter 10 “Accessories and spare parts”.

Check the instrument for any damage that may have been caused.

In the event of any damage, do not commission the instrument and contact the manufacturer immediately.

EN



CAUTION!


Damage to property due to electrostatic discharge (ESD)

When working on open circuits (PCBs) there is a danger of damaging sensitive electronic components through electrostatic discharge.

- ▶ When the battery compartment is open, e.g. when changing the battery, sufficient ESD protection must be ensured.
- ▶ Do not touch PCBs and electrical components.
- ▶ Before removing the plastic cover, touch any part of an adjacent grounded metal object (e.g. radiator, pipelines) (static charges are dissipated from the body).
- ▶ Avoid contact between the electronics and clothing.

5.1 Commissioning procedure

1. Fasten the instrument at the desired mounting location, see chapter 5.2 “Instrument mounting”
2. Connect the power supply and the input and output signals, see chapter 5.3 “Electrical connection” and chapter 5.3.2 “Battery power (A2G-500)”
3. Carry out a zero point setting (depending on instrument variant, see chapter 5.4.5 “Zero point setting”).
4. Connect measuring hoses, see following graphic




5. The instrument is ready for configuration, see chapter 5.4 “Operation / Instrument setting”

5. Commissioning and operation


5.2 Instrument mounting

The instrument can be mounted both vertically and horizontally. Measurement accuracy and IP ingress protection may deviate with cable entry from side. The specifications are based on vertical mounting with cable entries at the bottom.

EN



1. Select a vibration-free mounting location



a. Fastening to a duct, wall or panel.

b. Fastening to a DIN rail (standard rail TS35 (35 x 7.5 mm) in accordance with DIN EN 60715)


2a. Separate the mounting plate from the case and use the screw holes as drilling templates.
(bore diameter 3.8 mm)

2b. Snap the instrument onto the DIN rail.

3a. Mount using the fastening screws supplied.

5. Commissioning and operation

Application-related connections




EN

5.3 Electrical connection



The instrument is designed to operate with safety extra-low voltages (SELV).

1. Remove the case cover



5. Commissioning and operation



In the instrument version with battery supply, the batteries can be inserted directly after opening the case cover, see chapter 5.3.2 “Battery power (A2G-500)”.

EN

2. Loosen the cable gland and feed through the cable or cables.
3. Tighten the cable gland.
4. Connect the wires, see chapter 5.3.1 “Connection diagram”.

Load: Current output = max. 500 Ω
Voltage output = Typ. 1 k Ω , min. 700 Ω



CE conformity requires shielded cables. These must be grounded at the opposite end (PLC or DDC). With Modbus cabling, the shield must be connected on both sides to the specially designated connection terminals (labelled Shield).

5.3.1 Connection diagram

Explanation of abbreviations in the diagram.

Abbreviation	Meaning
UB +	Operating voltage
GND	Ground
AO x	Output signal
P	Plus (2-wire)
M	Minus (2-wire)
A	Current output
V	Voltage output
A +(x)	Output signal A +
B -(x)	Output signal B -
Shield	Shield
INx +/-	Active input signal (positive/negative)
INx P/N	Passive input signal (positive/negative)
R-IN x	Relay input
R-NO x	Relay normally open


x = Numbering

40452360.02 04/2025 EN

5. Commissioning and operation


A2G-500 (single-range)

Output signal 4 ... 20 mA, 2-wire





EN

Output signal 0 ... 10 V, 4 ... 20 mA, 0 ... 5 V or 2 ... 10 V, 3-wire



A2G-500 / -520 / -540 (multi-range)


Output signal 0 ... 10 V, 4 ... 20 mA, 0 ... 5 V or 2 ... 10 V, 3-wire



Second output signal included


5. Commissioning and operation

4 x input signal (2x voltage input, 2x resistance input)




Maximum cable length <10 m
[<32,80 ft]


Modbus connection



Additional signal output



2 x relay switching output



5.3.2 Battery power (A2G-500)



DANGER!

Physical injuries and damage to property and the environment through lithium thionyl chloride batteries

Improper handling of lithium thionyl chloride batteries can lead to leakage or escape of vapourised electrolyte vapours and cause a fire or explosion.

- ▶ Do not open the batteries.
- ▶ Do not damage the batteries.
- ▶ The positive and negative connections must not be short-circuited with conductors.
- ▶ Do not reverse the polarity.
- ▶ Do not expose the batteries to excessive mechanical loads.
- ▶ Do not expose the batteries to water or condensation.
- ▶ Do not heat, solder or expose the batteries to fire.
- ▶ Do not use any rechargeable batteries.
- ▶ Do not continue to use batteries that have been dropped on the ground / damaged.

To ensure proper operation and the best possible performance, use only the batteries listed below:

Lithium thionyl chloride battery, model Saft LS 14500


The battery life is transmitted as a percentage value via LoRaWAN®. The standard battery life is 5.5 years. As standard, the following parameter settings are used:

- 0.5 second response time
- 1 hour LoRa® interval
- SF10 radio quality
- 23 °C [73.4 °F] Temperature

At values below 20 % of the expected remaining service life a battery change is recommended.

The battery charge/life can be read out as a percentage value via the “myWIKa wireless device” app and the NFC interface.

5. Commissioning and operation



Polarity of the batteries

Battery (top)	Left = - Right = +
Battery (bottom)	Left = + Right = -

5.4 Operation / Instrument setting

The instrument can be operated in different ways, depending on the specification.

Model	A2G-500	A2G-500 / -520 / -540		A2G-500
Specification	Single-Range	Multi-Range	Modbus RTU	Wireless (battery power)
WIKA app (NFC)				
Configure	✓	✓	✓	✓
Read	✓	✓	✓	✓
Button/Display	✗	O	O	✗
Button/LED	O ¹⁾	O ²⁾	✗	✗

1) Only 3-wire

2) Only A2G-500

✓: included

✗: not included

O: optional

5. Commissioning and operation



All setting options are available via the WIKA app (via NFC) and the buttons/display.

The setting options are limited if using the buttons/LED.

In the A2G-520 and A2G-540 versions without display, the buttons have no function.

EN

5.4.1 Operation via the WIKA app (NFC)

Via the “myWIKa wireless device” app and an NFC connection, the data transmission can be activated and deactivated through the mobile device. Furthermore, the instrument data and the current measured value can be read and the parameters set via the app.

“myWIKa wireless device” app



For iOS-based mobile phones (from iOS version 13), the app is available in the Apple Store under the link below:



For mobile phones with an Android operating system, the app is available in the Play Store via the link below:




5. Commissioning and operation

The structure of the WIKA app is explained below:

Select language: The language can be selected under “Settings” in the lower menu ribbon. The model A2G-5x0 instrument can then be selected on the start screen.


EN



Smartphone positioning for NFC data exchange:

The NFC antenna on the A2G-5x0 is installed on the printed circuit board (see figure below). The NFC antenna on the smartphone is installed differently by each manufacturer. It is always located on the back of the smartphone. It can be at the top, in the centre or at the bottom. It is therefore important to find the right position.


In addition, the NFC communication must be activated in the smartphone settings.



iPhone models from version 7 are NFC-capable.
With some Android devices, NFC is deactivated as standard.

5. Commissioning and operation

Hold the smartphone **right up against** the A2G-5x0. If the smartphone position is correct, "Communication active ..." appears on the A2G-5x0 display.



NFC

Communication
active...



EN

Data exchange successful

Data exchange failed, try again



Change smartphone position and
hold it right up against the A2G-5x0



5. Commissioning and operation

Menu structure

The menu structure, menu items and settings are explained below.

Some settings are self-explanatory, such as the pressure ranges, and are not explained further.


Depending on the instrument version and options, only the menu items that are available on the hardware side can be set in the instrument. The options that have not been ordered are greyed out and cannot be selected.

EN

Device home




Settings




5. Commissioning and operation

EN


- ① **Request data:**
Read data, specifications and settings of the instrument




- ② **Load settings:**
Upload saved parameter files (settings) from the smartphone to the instrument



- ③ **Dashboard:**
Display of current measured values




- ④ **Device Actions:**
- Zero point adjustment (manual and automatic, depending on hardware)
 - Manual connection of the instrument and the LoRaWAN® gateway
 - Transmission of current measured value via LoRaWAN®
 - Changing the battery (for battery-powered instruments)
 - Resetting the instrument to factory settings
- Factory setting password: 1234



40452360.02 04/2025 EN


5. Commissioning and operation

⑤ Device information: Instrument information




⑥ Settings: Parameter settings

- The orange dot indicates that change(s) have been made.
- ❗ A red, circled exclamation mark indicates that the entered value is not possible.



⑦ Write settings to the A2G: Sending changed specifications to the instrument



⑧ Share settings: Save changed, customised and current settings as a file on your smartphone or send them via e-mail, for example



5. Commissioning and operation

⑨ Lock function:

The password can be activated or deactivated and the password can be changed in the “General” settings, see chapter 5.4.1.8 “General”.

Factory setting password: 1234





EN

5. Commissioning and operation

Alphanumeric designation

In the respective menu items, in this example “Pressure”, the measured value can be described alphanumerically, e.g. “SUP FAN F4” (max. 13 characters).

In the “Measuring range” setting, individual measuring ranges can be selected in addition to the predefined measuring ranges, i.e. a customised measuring range can be defined.



5.4.1.1 Air flow (A2G-520 and A2G-540)

1. Select “Air flow”

Select unit for display and output signal

Flow unit: m^3/s , m^3/h , cfm , l/s

Formula unit: m/s , ft/min


2. Calculation formula (manufacturer’s formula):

Measurement of the air volume flow using the K-value of the ventilator

a) Select ventilator manufacturer

b) “Common probe” (measuring probe):

Use together with the model A2G-FM measuring probe



5. Commissioning and operation

EN

Manufacturers' calculation formulas

Calculation formula	Formula unit	Formula	K-value	Flow unit
Common Probe	m ³ /h; m ³ /s; l/s; cfm	$q = k * \sqrt{\Delta p}$	0.001 – 9999.99	m ³ /h; m ³ /s; l/s; cfm
	m/s; ft/min	$v = k * \sqrt{\Delta p}$	0.001 – 9999.99	m/s; ft/min
Rosenberg	m ³ /h	$q = k * \sqrt{\frac{2}{\rho} \Delta p}$	37.0 - 800.0	m ³ /h; m ³ /s; l/s; cfm
Comefri	m ³ /h	$q = k * \sqrt{\frac{2}{\rho} \Delta p}$	10.0 – 2000.0	m ³ /h; m ³ /s; l/s; cfm
Nicotra Gebhardt	m ³ /h	$q = k * \sqrt{\frac{2}{\rho} \Delta p}$	10.0 – 4700.0	m ³ /h; m ³ /s; l/s; cfm
Ziehl-Abegg	m ³ /h	$q = k * \sqrt{\Delta p}$	10.0 – 1500.0	m ³ /h; m ³ /s; l/s; cfm
ebm-papst	m ³ /h	$q = k * \sqrt{\Delta p}$	10.0 – 1500.0	m ³ /h; m ³ /s; l/s; cfm
Fläkt Woods	m ³ /s	$q = \frac{1}{k} * \sqrt{\Delta p}$	0.3 – 99.0	m ³ /h; m ³ /s; l/s; cfm

Legend:

q = Air flow

k = K-value

ρ = Air density at standard conditions

Δp = Differential pressure

v = Velocity

3. Only for “Common Probe”

The unit for the formula and the flow must be selected identically. Subsequently, the K-value is entered.

4. Select K value: between 0.001 ... 9999.99

The ventilator- or probe-specific K value must be entered. The K value can be found on the data sheet / product label of the respective ventilator or installation part.

5. Select the maximum air flow (upper measuring range limit) for scaling the output signal

In the “Scale air flow output” menu item, enter the maximum air flow depending on the selected measuring instrument variant.

5. Commissioning and operation

6. Noise cancelling

The noise cancelling function can be used to suppress the smallest air movements when the ventilator is switched off.

7. Temperature and/or pressure compensation

Optionally, a temperature and/or pressure compensation can be carried out.

Temperature compensation:

If a temperature signal $T(x)$ is connected to the inputs of the instrument, see chapter 5.4.1.3 „Input“, temperature compensation can be activated in the air flow. The pressure difference for the air flow is adjusted according to the temperature applied in relation to the standard environment.

$$q_{T-komp.} = q * \sqrt{\frac{T(x)}{T_{ref}}} \quad T_{ref} = 20 \text{ }^{\circ}\text{C}$$

q_T = Air flow, temperature compensated

T_{ref} = Temperature reference = 20 °C [68 °F]

Pressure compensation:

There are two options for pressure compensation in the air flow menu: altitude correction and an external pressure signal. With altitude correction, the altitude of the measuring location above sea level can be entered and the pressure difference for the air flow is compensated according to the isothermal barometric altitude formula. If an external pressure signal, e.g. for the external pressure, is connected to an analogue input, see chapter 5.4.1.3 „Input“ of the instrument, pressure compensation is performed in relation to the standard environment.

$$q_{p-komp.} = q * \sqrt{\frac{p_{ref}}{p(h)}}$$

With altitude correction:

$$p(h) = p_{ref} \cdot e^{-\frac{g \cdot \rho_0}{p_{ref}} \cdot h}$$

$$p_{ref} = 101325 \text{ [Pa]}$$

$$g = 9.80665 \left[\frac{\text{m}}{\text{s}^2} \right]$$

$$\rho_0 = 1.2041 \left[\frac{\text{kg}}{\text{m}^3} \right]$$

$$h = \text{altitude [m]}$$

The values with the index 0 refer to the reference values of the ICAO standard atmosphere and not to the laboratory conditions.

5. Commissioning and operation

Or ambient pressure sensor: $p(h)$ = value from external ambient pressure sensor

q_p = Pressure compensated

p_{ref} = Pressure reference

5.4.1.2 PID controller (A2G-540)

The controller output is calculated by the instrument using:

$$u = 0.5 + K_P \cdot e + K_I \cdot \int e(t)dt + K_D \cdot \frac{d e(t)}{dt}$$

u = Manipulated variable (controller output)

e = Deviation from the set point

$K_P \cdot e$ = P = Proportional action

$K_I \cdot \int e(t)dt$ = I = Integral action


$K_D \cdot \frac{de(t)}{dt}$ = D = Derivative component

Possible controllers are P; PI; PD and PID controllers, controlled via the de-/activation of the P, I and D values.

1. Select “Flow” or “Pressure” under Controller

Set the setpoint:

- The source can be selected in the “Setpoint source” menu, e.g. external potentiometer.
- If the setpoint source “Manual setpoint” is selected, the value can be entered numerically in this menu.



5. Commissioning and operation

2. Select proportional band depending on the specification

Enter it in the "P-Value"

P in the respective pressure/flow unit for K_P value, reciprocal of P value ($K_P = 1/P$)

3. Select reset time

Enter it in "I-Value"

Tn in seconds for I-Value

4. Select derivative time

Enter in "D-Value"

Tv in seconds for D-Value

5. Noise cancelling


Enter the desired value for activation

6. Select forced output value

1. Deactivate or activate in the "Output forcing" menu

2. Manually enter the value between 0 and 1 (up to 5 decimal places possible) in the "Forced output value" menu

Example: P controller



Explanations of P controller figure:

- The user input for the P value as a proportional band in the range 0 ... 999999 (flow/pressure)
- The proportional band is defined as the reciprocal of K_P

5. Commissioning and operation

- The controller output is calculated using
$$u = 0.5 + K_P \cdot e$$
- Example: $P = 400 \text{ Pa} \Rightarrow K_P = 1/400\text{Pa}$
With a deviation of 100 Pa from the set point, the result is
$$u = 0.5 + 1/400 \text{ Pa} \cdot 100 \text{ Pa} = 0.75$$

Scaled to the $0 \dots 10 \text{ V}$ output, the controller outputs 7.5 V for this deviation

EN

5.4.1.3 Input

The active and passive inputs can be set in the “Input” setting.

The numerical value for the scaling must be entered in the “End point” and “Zero point” menu items.

The 'Input' settings menu is divided into two sections: 'Active input' and 'Passive input'. The 'Active input' section shows settings for 'Active IN 1' and 'Active IN 2'. For 'Active IN 1', the signal type is '0...10V', conversion unit is 'V', designation is empty, zero point is '0', and end point is '10'. The 'Passive input' section shows settings for 'Passive IN 3' and 'Passive IN 4'. For 'Passive IN 3', the signal type is 'Pt1000', conversion unit is '°C', and designation is empty. The same settings are shown for 'Passive IN 4'. In both sections, the 'Zero point [V]' and 'End point [V]' fields are highlighted with red boxes.

5.4.1.4 Output (relay)

Select Relay in “Output” menu

a. Select signal source

b. Select relay type


Select the desired type in the “Type” setting

- Normally open
- Normally closed
- Force open
- Force closed

The 'Output' settings menu shows two tabs: 'Analog output' and 'Relay'. The 'Relay' tab is selected. It shows settings for 'Relay 1' and 'Relay 2'. For 'Relay 1', the signal source is 'Pressure', type is 'Normally open', start 1 is '100', end 2 is '400', and hysteresis is '5'. For 'Relay 2', the signal source is 'None', type is 'Normally open', start 1 is '-100', end 2 is '100', and hysteresis is '0'. The 'Relay' tab is highlighted with a blue bar.

5. Commissioning and operation

EN



c. Select switch points

Enter the start and end value ("Zero value" and "End value") in the "Start 1" and "End 2" settings.

d. Enter the desired value in the "Hysteresis" setting

5.4.1.5 Wireless (LPWAN)

Integration of the radio network

The energy mode of the A2G-5x0 defines the communication class for LoRaWAN®.

A2G-5x0: LoRaWAN® class A

The A2G-5x0 is configured for "over-the-air" activation. To go through the activation procedure, the instrument is already preconfigured with specific LoRaWAN® parameters. In accordance with the LoRaWAN® specifications 1.1.0, the instrument is equipped with the following parameters:

- A device identifier (DevEUI)
- An application identifier (AppEUI)
- An application key (AppKey)

This information can be found on the enclosed label and must be communicated to the network server so that it can activate the instrument and communicate with the instrument. Commissioning the instrument is a specific process that must be carried out with a LoRaWAN® network service provider. Further information can be obtained from the network service provider.

Connection procedure

When switched on, the sensor starts a LoRaWAN® connection sequence (1 attempt and 1 retry 3 minutes later if the first attempt was unsuccessful).

40452360.02 04/2025 EN

5. Commissioning and operation

In the event of a failure, the sensor goes into sleep mode for a random period of time and then starts a new connection sequence.

The sleep mode between 2 connection sequences is defined as follows:

- 10 ... 15 min, first time
- 55 ... 60 min, second time
- 3 h 55 min ... 4 h, the subsequent times

The last time period is retained indefinitely until the connection is successful or is restarted.

LoRaWAN® LED

The LED is located on the LoRaWAN® board. The meanings are explained in the table below.

LED colour	Mode	Meaning
Cyan	Blinks once	LoRaWAN® join procedure successful
Red	Blinks twice	LoRaWAN® data transmission not successful (confirmed message)
Green	Blinks twice	LoRaWAN® data transmission successful (confirmed message)

EN

5. Commissioning and operation



WIKa app settings

- The “Manual LoRaWAN® JoinRequest” command can be started under “Device Actions”. The LED flashes once in cyan = the device is connected to the LoRaWAN® gateway.
- In the WIKa app under “Settings” - “Wireless”:
 - Select the transmission protocol.
 - Set the desired transmission rate (from 15 min ... 24 h).



If the instrument is connected to a TTN Community Network (LoRa), compliance with the Fair Use Policy is required. This means that a LoRa message may only be sent every 2 hours.

- Activate the radio module. When the radio module is switched off, the instrument does not send any data.



With a customer-specific integration, the payload integration must be carried out by the customer in accordance with the special documentation “Special documentation for LoRaWAN® communication specifications, model A2G-500 / A2G-520 / A2G-540”. The special documentation can be found on the WIKa website.




To successfully complete the cloud integration, an instrument restart is mandatory after the onboarding process (disconnect from the power supply for at least 5 seconds). The measured values only appear in the cloud after the instrument is restarted.

5. Commissioning and operation

5.4.1.6 Bus (Modbus®)

In the “Bus” menu, the Modbus® settings can be made.



EN



When the termination is switched on, the terminating resistor is activated. The instrument with the termination must be permanently energised.



The modbus® register is required for the modbus® connection. The register is included in the separate “special documentation” (article number 40455191). The special documentation can be found on the WIKA website.

5. Commissioning and operation


5.4.1.7 Display

In the “Display” menu item under “Dashboard”, you can select how many and which measured values (Value 1-4) are visible and whether the relays should be shown on the instrument display.

Start display example:

1 x value, relay


SUP FILTER F4 [Pa]	
250	
R1:open	R2:closed
SP1: 200Pa SP2: 450Pa	SP1: 2000m3/h SP2: 6500m3/h



The colour change of the set limit values can be defined under “Alarm”. The displayed value can take on up to three different colours. E.g., in the case of a supply air filter, the normal value can be displayed in WIKA blue, the pre-alarm in yellow and the alarm in red. The colours can be freely-selected using RGB.

e.g. RGB: 031, 000, 000 = red

SUP FILTER F4 [Pa]	
459	
R1:open	R2:closed
SP1: 200Pa SP2: 450Pa	SP1: 2000m3/h SP2: 6500m3/h




5. Commissioning and operation

EN

5.4.1.8 General

In the “General” menu:

- The confirmation menu for operation with button/display can be activated or deactivated.
- The password function can be enabled or disabled and a password set.
- The background, foreground and title colour of the display can be defined on a customer-specific basis.



5.4.2 Operation via buttons/display



The buttons have no function with instrument versions without a display and without LEDs.




When using the buttons, ESD protective measures must be taken into account.

5. Commissioning and operation

Control element





The following buttons are used to control the operating menu on the display:

- Select
 - Select or call up menu = press briefly
 - Confirm or exit = press and hold for 1 s
- Up/Down
 - Navigate through menu (upwards/downwards)



The “Select” function may vary depending on the menu level.
Note the legend on the respective display.

Display examples



5.4.3 Operation via buttons/LED

Default settings can be configured as follows using 3 buttons and 7 LEDs:

- The “Up” and “Down” buttons are used to navigate between the LEDs.
- The LED is switched on or off with the “ON/OFF” button.

To show which LED is selected, this LED flashes at a defined regularity.

Switched on = Slow flashing (0.5 Hz = 1 second on and 1 second off)

Switched off = Normal flashing (1.67 Hz = 300 milliseconds on and 300 milliseconds off)

No default setting = Fast flashing (10 Hz = 50 milliseconds on and 50 milliseconds off)


5. Commissioning and operation



The WIKA app can be used to configure other settings that cannot be selected as default settings. When no default settings are set, LED 7 is lit permanently.

Description of the LEDs (default settings)

- LED 1: Pressure unit
 - ON = Pa
 - OFF = inWC
- LED 2, 3 and 4: Measuring range (displayed in binary) (only with multi-range instruments)
- LED 5: Output signal
 - ON = 0 ... 10 V
 - OFF = 4 ... 20 mA
- LED 6: Response time
 - ON = 4 s
 - OFF = 0.8 s
- LED 7: Zero point calibration
 - OFF = Not active
 - Blinks = Active



If LED 7 is lit continuously, one or more non-default setting(s) have been configured with the WIKA app.

Binary code			Operating range ±250 Pa	Operating range ±2,500 Pa	Operating range ±7,000 Pa	Operating range ±12,000 Pa
LED 4	LED 3	LED 2				
Off	Off	Off	0 ... 25	0 ... 250	0 ... 1,000	0 ... 5,000
Off	Off	On	0 ... 50	0 ... 500	0 ... 1,500	0 ... 6,000
Off	On	Off	0 ... 100	0 ... 1,000	0 ... 2,000	0 ... 7,000
Off	On	On	0 ... 250	0 ... 1,500	0 ... 2,500	0 ... 7,500
On	Off	Off	-25 ... +25	0 ... 2,000	0 ... 3,000	0 ... 8,000
On	Off	On	-50 ... +50	0 ... 2,500	0 ... 4,000	0 ... 9,000
On	On	Off	-100 ... +100	-500 ... +500	0 ... 5,000	0 ... 10,000
On	On	On	-150 ... +150	-1,000 ... +1,000	0 ... 7,000	0 ... 12,000


LED 1 has the following functionality:

- 10 x flashing with a frequency of 1.67 Hz (300 ms on and 300 ms off) on start
- Constantly on during NFC communication
- Flashing in the event of an error

5. Commissioning and operation

5.4.4 Setting the response time

The response time affects how fast the transmitter reacts to changes in the pressure conditions in the system. The response time defines the time the measuring instrument needs in order to reach 63 % of the measured value. With unstable pressure conditions, a longer response time must be selected.




Example:

Selected response time: 4.0 seconds

Result: The output signal reaches a new value in 20 seconds (response time * 5)

In the “Pressure” menu, the response time can be set from 0.5 to 20 seconds.




5.4.5 Zero point setting

Manual zero point setting

1. Remove both hoses from the pressure connections $\oplus$ and $\ominus$.
2. The zero setting can be started via the WIKA app under “Device Actions” for all instrument versions. For multi-range instruments, the zero point setting can also be activated via the Select button or ON/OFF button (press for 5 seconds). Depending on the version, “Auto zero active ...” appears on the display or LED 7 starts to flash.
3. Wait until the information on the screen is no longer visible, LED 7 switches off or follow the instructions on the WIKA app.
4. Reconnect the hoses to the pressure connections.
5. Do not change the location of the sensor.

5. Commissioning and operation

EN




In normal operation, we recommend that a zero point calibration is carried out every 12 months.

Automatic zero point setting

The automatic zero point setting makes the instrument maintenance-free. The A2G-5x0 corrects the zero point at a defined interval and thus prevents any zero-point drift in the piezoresistive sensor element.

During the zero point setting the display and output value remains at the last measured value. Automatic zero point setting takes 5 seconds. The interval can be set either in the WIKA app, in the “Pressure” menu under “Autozero”, or on the display in the “Auto zero” menu.



40452360.02 04/2025 EN

6. Faults

6. Faults

Personnel: skilled electrical personnel

Tools: voltage tester, screwdriver

EN



If faults cannot be eliminated by means of the listed measures, the instrument must be taken out of operation immediately.

- Contact the manufacturer.
- If a return is needed, please follow the instructions given in chapter 8.2 "Return".



For contact details, see chapter 1 "General information" or the back page of the operating instructions.

Faults	Causes	Measures
Connection to the IIoT platform is not successful	Login credentials lost	Contact customer service
	Incorrect login credentials	Check using the supplied login credentials
	Customer firewall blocks interfaces	Contact the person responsible for infrastructure
	Instrument is outside the range of the gateway	Observe instructions in accordance with the operating instructions
	Faulty commissioning or improper, unsuitable installation location	Observe instructions in accordance with the operating instructions
Individual measured value not transmitted (LPWAN)	Collision in the data transmission	Unavoidable! Adaptation of infrastructure possible
NFC transmission does not work	NFC is switched off on the smartphone	Switch on NFC on the smartphone
	Smartphone is not held against the instrument in the right place	Hold the smartphone against the NFC sensor
		Hold the smartphone at the top or at the side of the instrument
Constant output signal upon change in pressure	Incorrect measuring range	Set the correct measuring range
	Incorrect signal source for the output	Set the correct output signal source

40452360.02 04/2025 EN

6. Faults

EN

Faults	Causes	Measures
Erroneous analogue input/output signal	Cable not properly connected	Check the cable connections
	Instrument set incorrectly	Check the settings
Instrument leaking	Cover not mounted correctly	Mount the case cover correctly
	Seal is missing/defective	Replace/mount the seal
	Cable gland / blind plug not mounted correctly	Mount the cable gland /blind plug correctly

Display error messages

Code	Name	Cause	Measure
0	ERROR_NONE	No error	No error
1	ERROR_EEPROM_NO_RESPONSE	■ ST25: Writing to RAM failed ■ ST25: Reading RAM failed	■ Follow the instructions on the screen ■ Restart the instrument if this occurs repeatedly
2	ERROR_EEPROM_INVALID_MAGICNR	Invalid user settings identified in the memory.	Default user settings are loaded automatically.
3	ERROR_FACTORY_INVALID_MAGICNR	Invalid factory options identified in the memory.	Default factory options are loaded automatically.
4	ERROR_FACTORY_ERASE_FAILED	Failed to delete the factory options from the flash memory.	Contact the manufacturer
5	ERROR_FACTORY_WRITE_FAILED	Failed to write the factory options from the flash memory.	Contact the manufacturer
6	ERROR_PRESSENS_TIMEOUT	This error appears if the pressure sensor does not display any new measurement data within a timeout.	■ Follow the instructions on the screen ■ Restart the instrument if this occurs several times
7	ERROR_PRESSENS_NO_RESPONSE	No response received from the pressure sensor.	Follow the instructions on the screen
8	ERROR_ADC_CONVERSION_FAILED	This error appears if the analogue digital conversion takes too long. (>10 ms).	■ Follow the instructions on the screen ■ Restart the instrument if this occurs several times

6. Faults

EN

Code	Name	Cause	Measure
9	ERROR_AOUT1_COMMON_MODE_OVER_RANGE_FAULT	This error appears if the analogue output AOUT1 displays the common-mode over-range error.	Check or connect signal at AOUT1
10	ERROR_AOUT1_LOAD_FAULT	This error appears if the analogue output AOUT1 displays the load fault error.	Check or connect signal at AOUT1
11	ERROR_AOUT1_OVER_TEMPERATURE_WARNING	This error appears if the analogue output AOUT1 displays the over-temperature error.	Instrument is too warm, leave it to cool
12	ERROR_AOUT2_COMMON_MODE_OVER_RANGE_FAULT	This error appears if the analogue output AOUT2 displays the common-mode over-range error.	Check or connect signal at AOUT2
13	ERROR_AOUT2_LOAD_FAULT	This error appears if the analogue output AOUT2 displays the load fault error.	Check or connect signal at AOUT2
14	ERROR_AOUT2_OVER_TEMPERATURE_WARNING	This error appears if the analogue output AOUT2 displays the over-temperature error.	Instrument is too warm, leave it to cool
15	ERROR_CONFIG_INVALID	Invalid signal source specified.	Restart the instrument if this occurs repeatedly
16	ERROR_PRESSENS_ZERO_POINT_OFFSET_OVER_RANGE_FAULT	Zero point adjustment Offset too large	Defective valve with auto-zero or pressure connected without auto-zero valve

7. Maintenance and cleaning

Personnel: skilled electrical personnel

Tools: voltage tester, screwdriver



For contact details, see chapter 1 “General information” or the back page of the operating instructions.

EN

7.1 Maintenance

The instrument version with automatic zero point adjustment is maintenance-free.

For the instrument version with manual zero point adjustment, this must be carried out regularly as described in chapter 5.4.5 “Zero point setting”.

Repairs must only be carried out by the manufacturer.

This does not apply to the battery replacement.

Only use original parts, see chapter 10 “Accessories and spare parts”.

7.2 Changing the batteries (only with instruments with battery operation)

For battery replacement observe the following:

- Only change the batteries in a dry environment.
- Do not use any rechargeable batteries.
- Only use approved batteries, see chapter 5.3.2 “Battery power (A2G-500)”.



During longer times of inactivity, remove the batteries from the instrument.

7.3 Cleaning



CAUTION!

Damage to property due to improper cleaning

Improper cleaning may lead to damage to the instrument.

- ▶ Do not use any aggressive cleaning agents.
- ▶ Do not use any hard or pointed objects for cleaning.
- ▶ Do not use any abrasive cloths or sponges.

7. Maintenance ... / 8. Dismounting, return and disposal

1. Before cleaning, correctly disconnect the instrument from the pressure supply, switch it off and disconnect it from the mains.
2. Clean the instrument with a moist cloth.
Electrical connections must not come into contact with moisture.
3. Clean the dismantled instrument, in order to protect persons and the environment from exposure to residual media.

EN


8. Dismounting, return and disposal

Personnel: skilled electrical personnel

Tools: voltage tester, screwdriver

8.1 Dismounting

Use a slotted screwdriver to separate the instrument from the mounting plate at the markings as shown in the following image.



8.2 Return

Strictly observe the following when shipping the instrument:

- All instruments delivered to WIKA must be free from any kind of hazardous substances (acids, bases, solutions, etc.) and must therefore be cleaned before being returned, see chapter 7.3 “Cleaning”.
- When returning the instrument, use the original packaging or a suitable transport packaging.

40452360.02 04/2025 EN

8. Dismounting, return and disposal



With hazardous substances, include the material safety data sheet for the corresponding medium.

EN

Instruments with lithium-ion rechargeable batteries or lithium-metal batteries

The lithium-ion rechargeable batteries or lithium-metal batteries included are subject to the requirements of the dangerous goods law. Special requirements for packaging and marking must be observed when shipping. A dangerous goods expert must be consulted when preparing the package. Do not send any damaged or defective rechargeable batteries. Mask open contacts and pack the rechargeable battery so that it does not move in the packaging and also prevents short-circuits. Observe the different dangerous goods requirements relative to the respective modes of transport and any other national regulations.

To avoid damage:

1. Wrap the instrument in an anti-static plastic film.
2. Place the instrument, along with the shock-absorbent material, in the packaging.
3. If possible, place a bag, containing a desiccant, inside the packaging.
4. Label the shipment as carriage of a highly sensitive measuring instrument.



Information on returns can be found under the heading “Service” on our local website (product return form).

8.3 Disposal

Incorrect disposal can put the environment at risk.

Dispose of instrument components and packaging materials in an environmentally compatible way and in accordance with the country-specific waste disposal regulations.

Disposal of electrical appliances with non-permanently installed batteries



This instrument is labelled in accordance with the EU Waste Electrical and Electronic Equipment (WEEE) directive. This instrument must not be disposed of with household waste.

- ▶ Hand in old instruments for environmentally friendly disposal at a designated collection point for the disposal of electrical and electronic devices.
- ▶ Ensure a proper disposal in accordance with national regulations.
- ▶ Observe the currently applicable regulations.
- ▶ Remove non-permanently installed batteries from the instrument and dispose of them separately.

8. Dismounting, return and disposal

Disposal of batteries



WARNING!

Damage to the environment and health due to incorrect disposal of batteries

Batteries contain pollutants such as heavy metals, which are harmful to the environment and health if not disposed of properly.

- ▶ Do not dispose of batteries with household waste.
- ▶ Ensure a proper disposal in accordance with national regulations.
- ▶ Observe the currently applicable regulations.
- ▶ Hand in used batteries for environmentally friendly disposal at retail outlets or appropriate collection points in accordance with national or local regulations.

If possible, completely discharge the batteries before disposal and isolate contacts to prevent short-circuits.

EN

9. Specifications

9. Specifications

9.1 Specifications A2G-5x0

Basic information	
Process connection ¹⁾	2 x connecting nozzle Ø 4.5 mm [0.177 in] / Ø 7.5 mm [0.295 in] For hoses with inner diameter 4 ... 6 mm [0.157 ... 0.236 in]
Case ¹⁾	PC GF20
Cover with insert sheet ¹⁾	PC, transparent; ABS, RAL 9010
Mounting plate ¹⁾	PC, RAL 7035
Weight	260 ... 320 g [0.57 ... 0.70 lb] (depending on version)

EN

1) Flame retardant according to UL94-V2

Measuring ranges ¹⁾ in Pa

Operating range ²⁾ ±250 Pa	Operating range ²⁾ ±2,500 Pa	Operating range ²⁾ ±7,000 Pa	Operating range ²⁾ ±12,000 Pa
-25 ... +25	-500 ... +500	0 ... 1,000	0 ... 5,000
-50 ... +50	-1,000 ... +1,000	0 ... 1,500	0 ... 6,000
-100 ... +100	0 ... 250	0 ... 2,000	0 ... 7,000
-150 ... +150	0 ... 500	0 ... 2,500	0 ... 7,500
0 ... 25	0 ... 1,000	0 ... 3,000	0 ... 8,000
0 ... 50	0 ... 1,500	0 ... 4,000	0 ... 9,000
0 ... 100	0 ... 2,000	0 ... 5,000	0 ... 10,000
0 ... 250	0 ... 2,500	0 ... 7,000	0 ... 12,000

- 1) Settable via instrument menu (display), HMI (buttons, LEDs) or WIKA app (NFC) or preset with single-range instruments
- 2) Defined with model code: Selection of an operating range (with several measuring ranges) or an individual measuring range

9. Specifications

Measuring ranges ¹⁾ in inWC

Operating range ²⁾ ±1 inWC	Operating range ²⁾ ±10 inWC	Operating range ²⁾ ±28 inWC	Operating range ²⁾ ±48 inWC
-0.1 ... +0.1	-2 ... +2	0 ... 4	0 ... 20
-0.2 ... +0.2	-4 ... +4	0 ... 6	0 ... 24
-0.4 ... +0.4	0 ... 1	0 ... 8	0 ... 28
-0.6 ... +0.6	0 ... 2	0 ... 10	0 ... 30
0 ... 0.1	0 ... 4	0 ... 12	0 ... 32
0 ... 0.2	0 ... 6	0 ... 16	0 ... 36
0 ... 0.4	0 ... 8	0 ... 20	0 ... 40
0 ... 1	0 ... 10	0 ... 28	0 ... 48

1) Settable via instrument menu (display), HMI (buttons, LEDs) or WIKA app (NFC) or preset with single-range instruments

2) Defined with model code: Selection of an operating range (with several measuring ranges) or an individual measuring range

Pressure type / Measuring element / Digital display

Pressure type	Differential pressure
Unit	
Differential pressure	■ Pa ■ kPa ■ mbar ■ mmWC ■ inWC
Air flow ¹⁾	■ l/s ■ m³/s ■ m³/h ■ cfm
Air velocity ¹⁾	■ m/s ■ fpm
Operating pressure	■ ±250 ... ±2,500 Pa [±1 ... ±10 inWC]: 10 kPa [40 inWC] ■ ±7,000 ... ±12,000 Pa [±28 ... ±48 inWC]: 100 kPa [400 inWC]
Measuring element	Piezo measuring cell
Digital display	
Display resolution	240 x 320 dpi
Type of display	2" TFT colour display

9. Specifications

EN

Pressure type / Measuring element / Digital display

Dashboard function	Indication of up to four measured values and two relay states possible; automatic size scaling depending on the number of measured values displayed. Indicated values can be given any alphanumeric designation.
Colour configuration	Range of setting possibilities; colours definable via RGB code.
Alarm function	Definition of two limit values possible; when these limit values are reached, a colour change can be defined for each (e.g. "Warning" and "Alarm")

- 1) The units of air flow and air velocity are only available on the A2G-520 and A2G-540.

Accuracy specifications

Accuracy ¹⁾	0.50 % FS ²⁾			
Operating range	±250 Pa [±1 inWC]	±2,500 Pa [±10 inWC]	±7,000 Pa [±28 inWC]	±12,000 Pa [±48 inWC]
Temperature range				
Compensated ³⁾ TC ⁴⁾ offset (% FS/K)	0.015 %	0.01 %	0.008 %	0.005 %
Uncompensated ⁵⁾ TC ⁴⁾ offset (% FS/K)	0.025 %	0.02 %	0.02 %	0.01 %
Compensated ³⁾ TC ⁴⁾ span (% FS/K)	0.02 %	0.02 %	0.01 %	0.01 %
Uncompensated ⁵⁾ TC ⁴⁾ span (% FS/K)	0.03 %	0.02 %	0.02 %	0.02 %
Zero point setting	■ Manually via push button on the PCB ■ "myWIKa wireless device" app ■ Automatic ⁶⁾			

- 1) At 23 °C [73.4 °F], relative to reference measuring instrument
Output wiring: 200 Ω in series for current output, 1000 Ω in parallel for voltage output
2) E.g. 0.50 % of ±250 Pa = 1.25 Pa
3) 0 ... 50 °C [32 ... 122 °F]
4) Temperature coefficient
5) -40 ... 0 °C [-40 ... +32 °F] / >50 °C [>122 °F]
6) Recommended for measuring ranges ≤ 250 Pa [≤ 1 inWC]

Output and input signal

Output signals	■ 0 ... 10 V, 0 ... 5 V or 2 ... 10 V, 3-wire ■ 4 ... 20 mA, 2- or 3-wire ■ Relay ■ Modbus® ■ LoRaWAN®
Number of analogue outputs	■ 1 x ■ 2 x

40452360.02 04/2025 EN

9. Specifications

EN

Output and input signal		
Number of relays	■ Without ■ 1 x (24 V / 0,5 A) ■ 2 x (24 V / 0,5 A)	
Switching function	■ NO (normally open) ■ NC (normally closed) ■ Force open, force closed	
Load	Current output	Max. 500 Ω
	Voltage output	Typ. 1 kΩ min. 700 Ω
Input signal		
2 x voltage input	■ DC 0 ... 10 V ■ DC 0 ... 5 V ■ DC 2 ... 10 V	
2 x resistance input	■ Pt1000 ■ Ni1000 ■ Ni1000-LG ■ NTC 10 kΩ (beta) = 3977 ■ Binary (on/off)	
Temperature range (resistance inputs)	-40 ... +120 °C [-40 ... +248 °F]	

Electrical connection	
Connection type	Spring-clip terminals
Wire cross-section	Max. 1.5 mm ²
Cable specification	Use shielded cables ¹⁾
Cable connection	■ Cable gland M20 x 1.5 with strain relief ■ Standard version with clamping area 4 ... 13 mm [0.15 ... 0.51 in] ■ Optional further sealing inserts available (see Accessories)
Reverse polarity protection	Integrated protection against reverse polarity and overvoltage resulting from wiring errors
Auxiliary power	
Supply voltage	■ AC 24 V ±10 % / 50 Hz / 60 Hz ²⁾ ■ DC 24 V ±10 %
Power consumption	■ 1.7 W at DC 24 V ■ 2.5 VA at AC 24 V
Electrical safety	Protection class III, safety extra-low voltage (SELV)

1) Instrument specifications are based on tests with shielded cables

2) Not permissible for 2-wire 4 ... 20 mA

9. Specifications

EN

Operating conditions	
Medium and ambient temperature range	■ -40 ... +60 °C [-40 ... +140 °F] ■ -10 ... +50 °C [14 ... 122 °F], with automatic zero point setting ■ -20 ... +60 °C [-4 ... +140 °F], with display
Storage temperature range	-40 ... +60 °C [-40 ... +140 °F]
Relative humidity, condensation	0 ... 95 % r. h., non-condensing
Permissible media	■ Air ■ Non-aggressive gases ■ Non-flammable gases
Mounting position	
Process connections	Lower mount ¹⁾
	Side mount ²⁾
Ingress protection of the complete instrument	IP65

1) Instrument specifications are based on tests with lower mount (reference position)

2) Deviation when side mount: +2 Pa; can be corrected via zero point function

NFC specification	
On-site interface	NFC (near field communication)
Standard	ISO/IEC 15693
Modulation	13.56 MHz

Modbus® communication	
Protocol	Modbus® via serial interface
Transfer mode	RTU
Interface	RS-485
Max. voltage input	■ ± 10 V without terminating resistor ■ ± 6 V with terminating resistor

9. Specifications

LoRaWAN® specification

LoRaWAN® specification	LoRa® 868 MHz EU
Version	1.0.3
Frequency range	863 ... 870 MHz
Transmission power	12 dBm
Range ¹⁾	≤ 10 km [≤ 6.2 mi]
Max. output power	14 dBm

1) The range depends on the topography. 10 km [6.2 mi] can be achieved in free field conditions and with a spreading factor of 12.

9.2 Approvals

Logo	Description	Region
	EU declaration of conformity	European Union
	EMC directive	
	RED - Radio Equipment Directive	
	RoHS directive	
	WEEE directive	

9.3 Certificates

Certificates


Certificates	■ Without ■ Measurement report per EN 837 ■ 2.2 test report per EN 10204 (e.g. state-of-the-art manufacturing, material proof, indication accuracy) ■ 3.1 inspection certificate per EN 10204 (e.g. material proof for wetted metal parts, indication accuracy, calibration certificate)
---------------------	---

- For approvals and certificates, see website
- Depending on the selected instrument version, the specification may deviate from the specifications listed here.
- The specifications in the order documentation are definitive.

For further specifications, see WIKA data sheets PE 88.05, PE 88.06 and PE 88.07 and order documentation.

9. Specifications

9.4 Dimensions in mm [in]



EN

10. Accessories and spare parts

10. Accessories and spare parts

Model	Description	Order number
	Measuring hoses	
	PVC hose, inner diameter 4 mm [0.16 in], roll at 25 m [82.02 ft]	40217841
	PVC hose, inner diameter 6 mm [0.24 in], roll at 25 m [82.02 ft]	40217850
	Silicone hose, inner diameter 4 mm [0.16 in], roll at 25 m [82.02 ft]	40217906
	Silicone hose, inner diameter 6 mm [0.24 in], roll at 25 m [82.02 ft]	40217914
	2 pcs. of duct connector for measuring hoses Ø 4 ... 6 mm [0.16 ... 0.24 in] and 4 pcs. of mounting screw	40217507
	Multiple sealing inserts (for M20 cable gland)	
	Sealing insert with 2 bores at Ø 4 mm [0.16 in]; PU 10 pcs.	40444284
	Sealing insert with 2 bores at Ø 6 mm [0.24 in]; PU 10 pcs.	40444285
	Sealing insert with 4 bores at Ø 5 mm [0.20 in]; PU 10 pcs.	40444286
	Battery AA 3.6 V, 2.6 Ah (only with battery version); PU 2 pcs.	40443947
	Sealing plugs for sealing insert with 4 bores à Ø 5 mm [0.20 in]; PU 50 pcs.	40454873

WIKA accessories can be found online at www.wika.com.

EU-Konformitätserklärung EU Declaration of Conformity



EN

Dokument Nr.: 40445841.01
Document No.:

Wir erklären in alleiniger Verantwortung, dass die mit CE gekennzeichneten Produkte
We declare under our sole responsibility that the CE marked products

Typenbezeichnung: A2G-500
Type Designation:

Beschreibung: Differenzdrucksensor für die Lüftungs- und
Description: Climatechnik
Differential pressure sensor for ventilation and air-conditioning

gemäß gültigem Datenblatt: PE 88.05
according to the valid data sheet:

mit den nachfolgenden relevanten Harmonisierungsvorschriften der Union übereinstimmen: Angewandte harmonisierte Normen:
are in conformity with the following relevant Union harmonisation legislation: Applied harmonized standards:

2011/65/EU gefährliche Stoffe (RoHS)
Hazardous Substances (RoHS)

EN IEC 63000:2018
Gesundheit und Sicherheit (Artikel 3 (1) a))
Protection of health and safety (Article 3 (1) (a))
EN 60730-1:2011
EN 60730-2-6:2016
EN 60730-2-15:2010
EN 62479:2010
Elektromagnetische Verträglichkeit (Artikel 3 (1) b))
Electromagnetic compatibility (Article 3 (1) b))
EN 301 489-1 V2.2.3
EN 301 489-3 V2.3.2

2014/53/EU Funkanlagen (RED)
Radio Equipment (RED)

stimmt auch überein mit/also complies with
EN 61326-1:2013
Effiziente Nutzung Frequenzspektrum (Artikel 3 (2))
Effective use of spectrum (Article 3 (2))
EN 300 220-1 V3.1.1
EN 300 220-2 V3.1.1
EN 300 330 V2.1.1

Unterzeichnet für und im Namen von / Signed for and on behalf of

WIKA Schweiz AG

Hitzkirch, 2023-07-28

Peter Barmettler, Technical Director
WIKA Schweiz AG

Peter Kung, Approval Representative
WIKA Schweiz AG

WIKA Schweiz AG
Industriestrasse 11
CH-6285 Hitzkirch

Tel. +41 (0) 41 919 72 72
Fax +41 (0) 41 919 72 72
E-Mail info@wika.ch

40452360.02 04/2025 EN

EU-Konformitätserklärung EU Declaration of Conformity



Dokument Nr.: 40445842.01
Document No.:

Wir erklären in alleiniger Verantwortung, dass die mit CE gekennzeichneten Produkte
We declare under our sole responsibility that the CE marked products

Typenbezeichnung: A2G-520
Type Designation:

Beschreibung: Differenzdruck-Volumenstromsensor für die Lüftungs- und Klimatechnik
Description: *Differential pressure air flow sensor for ventilation and air-conditioning*

gemäß gültigem Datenblatt: PE 88.06
according to the valid data sheet:

mit den nachfolgenden relevanten Harmonisierungsvorschriften der Union übereinstimmen:
are in conformity with the following relevant Union harmonisation legislation:

Angewandte harmonisierte Normen:
Applied harmonized standards:

2011/65/EU gefährliche Stoffe (RoHS)
Hazardous Substances (RoHS)

EN IEC 63000:2018
Gesundheit und Sicherheit (Artikel 3 (1) a)
Protection of health and safety (Article 3 (1) (a))
EN 60730-1:2011
EN 60730-2-6:2016
EN 60730-2-15:2010
EN 62479:2010
Elektromagnetische Verträglichkeit (Artikel 3 (1) b)
Electromagnetic compatibility (Article 3 (1) b))
EN 301 489-1 V2.3.3
EN 301 489-3 V2.3.2

2014/53/EU Funkanlagen (RED)
Radio Equipment (RED)

stimmt auch überein mit/also complies with
EN 61326-1:2013
Effiziente Nutzung Frequenzspektrum (Artikel 3 (2))
Effective use of spectrum (Article 3 (2))
EN 300 220-1 V3.1.1
EN 300 220-2 V3.1.1
EN 300 330 V2.1.1

Unterzeichnet für und im Namen von / Signed for and on behalf of

WIKAL Schweiz AG

Hitzkirch, 2023-07-28

Peter Barmettler, Technical Director
WIKAL Schweiz AG

Peter Kung, Approval Representative
WIKAL Schweiz AG

WIKAL Schweiz AG
Industriestrasse 11
CH-6285 Hitzkirch

Tel. +41 (0) 41 919 72 72
Fax +41 (0) 41 919 72 72
E-Mail info@wika.ch

EU-Konformitätserklärung EU Declaration of Conformity



Dokument Nr.: 40445843.01
Document No.:

Wir erklären in alleiniger Verantwortung, dass die mit CE gekennzeichneten Produkte
We declare under our sole responsibility that the CE marked products

Typenbezeichnung: A2G-540
Type Designation:

Beschreibung: Differenzdruck- und Volumenstromregler für die
Description: Differential pressure and air flow controller for
ventilation and air-conditioning

gemäß gültigem Datenblatt: PE 88.07
according to the valid data sheet:

mit den nachfolgenden relevanten Harmonisierungsvorschriften der Union übereinstimmen:
are in conformity with the following relevant Union harmonisation legislation:

Angewandte harmonisierte Normen:
Applied harmonized standards:

2011/65/EU gefährliche Stoffe (RoHS)
Hazardous Substances (RoHS)

EN IEC 63000:2018
Gesundheit und Sicherheit (Artikel 3 (1) a)
Protection of health and safety (Article 3 (1) (a))
EN 60730-1:2011
EN 60730-2-6:2016
EN 60730-2-15:2010
EN 62479:2010
Elektromagnetische Verträglichkeit (Artikel 3 (1) b)
Electromagnetic compatibility (Article 3 (1) b))
EN 301 489-1 V2.3.3
EN 301 489-3 V2.3.2

2014/53/EU Funkanlagen (RED)
Radio Equipment (RED)

stimmt auch überein mit/also complies with
EN 61326-1:2013

Effiziente Nutzung Frequenzspektrum (Artikel 3 (2))
Effective use of spectrum (Article 3 (2))
EN 300 220-1 V3.1.1
EN 300 220-2 V3.1.1
EN 300 330 V2.1.1

Unterzeichnet für und im Namen von / Signed for and on behalf of

WIKA Schweiz AG

Hitzkirch, 2023-07-28

Peter Barmettler, Technical Director
WIKA Schweiz AG

Peter Küng, Approval Representative
WIKA Schweiz AG

WIKA Schweiz AG
Industriestrasse 11
CH-6285 Hitzkirch

Tel. +41 (0) 41 919 72 72
Fax +41 (0) 41 919 72 72
E-Mail info@wika.ch

40452360.02 04/2025 EN

The LoRa® brand and the LoRa® logo are trademarks of Semtech Corporation.
LoRaWAN® is a trademark used under licence from LoRa-Alliance®.
Modbus® is a registered trademark of Schneider Electric.

WIKA subsidiaries worldwide can be found online at www.wika.com.



Importer for UK
WIKA Instruments Ltd
Unit 6 and 7 Goya Business park
The Moor Road
Sevenoaks
Kent
TN14 5GY



WIKA Alexander Wiegand SE & Co. KG
Alexander-Wiegand-Strasse 30
63911 Klingenberg • Germany
Tel. +49 9372 132-0
info@wika.de
www.wika.de